



DATE: 3/3/15

ADDENDUM # 1

BID NO. : 100071
OPENING: 3/20/15 3:00 PM
COMMODITY: Chromatographs, Gas

The following questions have been asked regarding bid 100071:

1. No specs on item #3 VOC EPA 524 method package.

Response: This is EPA 524 method for volatile organic carbon in drinking water. More specifically, the methods are pertinent to disinfectant by products such as Trihalomethanes.

2. Method is a purge and trap method, do you currently have a purge and trap apparatus that you will use or do you want it included?

Response: We have a purge and trap system that will interface with the new GCMS.

3. The bid does not contain any GC columns or other accessories?

Response: We have appropriate columns for this new method. No accessories required at this time.

4. When you refer to chemical resistance for the sipper, please elaborate on the types of samples i.e. acids,bases,organics,etc.

Response: The type of samples will be organic.

5. You mention validation software. Do you require 21CFR part 11?

Yes, the validation software is required.

6. Our system requires an external computer to control the UV/VIS Spectrophotometer, would this be something you could supply or would you purchase from us?

Response: No external computer will be purchased at this time.

Karol Chadwick
Assistant University Purchasing Agent