



DATE: 3/9/15

ADDENDUM#2

BID NO. : 100072
OPENING: 3/16/15 @3:00PM
COMMODITY: Lab Equipment

Please be advise that the following questions were asked as regards bid #100072 and the responses from the department follow.

Q1: Does the lab have an older Ultratrifuge now?If so what make and model?

A1:No, they do not have any Ultracentrifuges in the lab at this time

Q2: Would there be a trade -in for the old instrument? Or is this a shared unit with other labs/ departments

A2: No, there will not be a trade-in

Q3: Would there be any need for tubes for the applications in hand?

A3: Yes we will need tubes. We will need qty 1 (50 pack) of a polypropylene tube to fit the rotors. dimensions are 25x89mm, 32ml .

If you have already submitted a bid and need to make any changes , based on information within the Addendum , please submit a new bid response and indicate that this submission supersedes the prior.

Debra Lee
URI Assistant Purchasing Agent