

**THE
UNIVERSITY
OF RHODE ISLAND**

**DIVISION OF
ADMINISTRATION
AND FINANCE**

THINK BIG  WE DO™

PURCHASING DEPARTMENT

10 Tootell Road, Suite 3, Kingston, RI 02881 USA

p: 401.874.2171

f: 401.874.2306

uri.edu/purchasing



DATE: 1/25/16

ADDENDUM #1

BID NO.:100186
OPENING:2/3/16 @ 3:00pm
COMMODITY: Lab Equipment

Addendum #1 to Bid#100186 is being issued based on questions received by the URI Purchasing Dept. by the 1/21/16 deadline for questions. These questions and their corresponding answers are listed below:

Question 1: As regards the bid specification of "thermal conductivity: 0.001 to 1000W/MK"- Which is the upper requirement for thermal conductivity? Is it 600 or 1000 W/MK?

Answer 1: The upper requirement is 1000 W/MK at least. The higher the better, say 1500 or 2000 W/MK .

Question 2: This vendor is questioning the possibility of laser flash technique (as it relates to this bid) as an alternative?

Answer 2: We do not want a quote for laser flash technique for this bid at this time.

If you have already submitted a bid and need to make changes, based on the information within this Addendum, please submit a new bid response and indicate that this submission supersedes the prior.

Debra Lee
URI Assistant Purchasing Agent
Rev. 9-1-15