



Request for Quote

STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS
ONE CAPITOL HILL
PROVIDENCE RI 02908

CREATION DATE : 02-APR-12
BID NUMBER: 7449611
TITLE: MULTI GAS CALIBRATOR WITH FLOW CONTROLLERS & OZONE PHOTOMETER

BID CLOSING DATE AND TIME:02-MAY-2012 10:30:00

BUYER: McGurn, Cheryl A
PHONE #: N/A

**B
I
L
L
T
O**
DOA CONTROLLER
ONE CAPITOL HILL, 4TH FLOOR
SMITH ST
PROVIDENCE, RI 02908
US

**S
H
I
P
T
O**
DEM OFFICE OF AIR RESOURCES
235 PROMENADE ST, ROOM 230
PROVIDENCE, RI 02908
US

Requisition Number: 1254498
Note to Bidders: PER ATTACHED SPECIFICATIONS

VENDORS MUST SUPPLY SPECIFICATIONS WITH BID

Line	Description	Quantity	Unit	Unit Price	Total
1	SERIES 6103 MULTI-GAS CALIBRATOR WITH 3 FLOW CONTROLLERS & OZONE PHOTOMETER	1 00	Each		

Delivery: _____

Terms of Payment: _____

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer

Multi-gas calibrator with three flow controllers and ozone photometer

- The three flow controllers must be: 10 lpm zero air; 100 cc/min and 20 cc/min span, with program control of the span flow controllers, such that the software chooses the appropriate unit.
- The ozone photometer must be software-controlled, with feedback to the ozone generator for stabilization of the generator output. The photometer must accept input from internal and external ozone sources. The ozone generator must give stable output concentrations in the 20 to 400 ppb concentration range. It must have stabilization options which include photometer feedback and feedback from a sensor monitoring the ozone generator lamp.
- The reference air for the ozone photometer must have a solenoid valve upstream to eliminate zero-air consumption while the photometer is not in use.
- The calibrator must have at least three separate ports for span gas cylinders with software control of which port is accessed
- The calibrator must have internal, programmable software to control all aspects of the calibrator's functions. The software must allow for the storage and recall of more than 50 gas-concentration levels and for these to be assembled into at least 20 analytical sequences which can be programmed to execute automatically on any schedule devised by the user.

Contract Terms and Conditions

Table of Contents

Terms and ConditionsVI
BID STANDARD TERMS AND CONDITIONSVI
TERMS AND CONDITIONS FOR THIS BIDVI
CAMPAIGN FINANCE COMPLIANCEVI
PAYMENT AFTER FULL VENDOR PERFORMANCEVI
EQUAL OPPORTUNITY COMPLIANCEVI
VENDOR SPECIFICATIONSVI

Terms and Conditions

BID STANDARD TERMS AND CONDITIONS

TERMS AND CONDITIONS FOR THIS BID

CAMPAIGN FINANCE COMPLIANCE

EVERY PERSON OR BUSINESS ENTITY PROVIDING GOODS OR SERVICES AT A COST OF \$5000 CUMULATED VALUE IS REQUIRED TO FILE AN AFFIDAVIT REGARDING POLITICAL CAMPAIGN CONTRIBUTIONS WITH THE RI STATE BOARD OF ELECTIONS EVEN IF NO REPORTABLE CAMPAIGN CONTRIBUTIONS HAVE BEEN MADE. (RI GENERAL LAW 17-27) FORMS OBTAINED AT BOARD OF ELECTIONS, CAMPAIGN FINANCE DIVISION, 50 BRANCH AVENUE PROVIDENCE 02904 (401-222-2056)

ARRA SUPPLEMENTAL TERMS AND CONDITIONS

For contracts and sub-awards funded in whole or in part by the American Recovery and Reinvestment Act of 2009, Pub.L No. 111-5 and any amendments thereto, such contracts and sub-awards shall be subject to the Supplemental Terms and Conditions For Contracts and Sub-awards Funded in Whole or in Part by the American Recovery and Reinvestment Act of 2009, Pub L.No. 111-5 and any amendments thereto located on the Division of Purchases website at www.purchasing.ri.gov."

PAYMENT AFTER FULL VENDOR PERFORMANCE

PAYMENT SHALL NOT BE MADE UNTIL DELIVERY HAS BEEN MADE, OR SERVICES PERFORMED, IN FULL, AND ACCEPTED, UNLESS EXPRESSLY PROVIDED FOR HEREIN UPON THE RENDERING OF A PROPERLY SUBMITTED INVOICE.

EQUAL OPPORTUNITY COMPLIANCE

THIS PURCHASE ORDER IS AWARDED SUBJECT TO EQUAL OPPORTUNITY COMPLIANCE.

VENDOR SPECIFICATIONS

ALL VENDORS MUST INCLUDE SPECIFICATIONS WITH BID PROPOSAL (EVEN THOSE BIDDING BRAND SPECIFIED) FAILURE TO SUBMIT SPECIFICATIONS WITH BID PROPOSAL MAY RESULT IN DISQUALIFICATION OF BID ITEMS IN CATALOGS MUST BE CLEARLY MARKED AND PAGES TABBED