



Request for Quote

STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS
 ONE CAPITOL HILL
 PROVIDENCE RI 02908

CREATION DATE : 05-MAR-13
 BID NUMBER: 7461298
 TITLE: PORTABLE BLADDER SCANNER

 BID CLOSING DATE AND TIME:02-APR-2013 11:00:00

BUYER: McGurn, Cheryl A
 PHONE #: N/A

**B
I
L
L
T
O**
 DOA CONTROLLER
 ONE CAPITOL HILL, 4TH FLOOR
 SMITH ST
 PROVIDENCE, RI 02908
 US

**S
H
I
P
T
O**
 DOA-OLIS-INFORMATION PROCESSING
 ONE CAPITOL HILL, 2ND FLOOR
 PROVIDENCE, RI 02908
 US

Requisition Number: 1306249
 Note to Bidders: PER ATTACHED SPECIFICATIONS.

PLEASE PROVIDE VENDOR SPEC.

Line	Description	Quantity	Unit	Unit Price	Total
1	TOTAL COST FOR PORTABLE BLADDER SCANNER PRICE MUST INCLUDE: PORTABLE ROLLING CART ON SITE IN-SERVICE TRIANING FOR BLADDER SCANNER BLADDER SCANNER TRAINING PACK/VIDEO (ACUTE) (ENGLISH)	1.00	Each		

Delivery: _____

Terms of Payment: _____

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer

PURCHASE ORDER STANDARD TERMS AND CONDITIONS

TERMS AND CONDITIONS FOR THIS PURCHASE ORDER:

TERMS AND CONDITIONS FOR THIS BID

VENDOR SPECIFICATIONS

ALL VENDORS MUST INCLUDE SPECIFICATIONS WITH BID PROPOSAL (EVEN THOSE BIDDING BRAND SPECIFIED). FAILURE TO SUBMIT SPECIFICATIONS WITH BID PROPOSAL MAY RESULT IN DISQUALIFICATION OF BID. ITEMS IN CATALOGS MUST BE CLEARLY MARKED AND PAGES TABBED.

INSURANCE REQUIREMENTS

AN INSURANCE CERTIFICATE IN COMPLIANCE WITH PROVISIONS OF ITEM 31 (INSURANCE) OF THE GENERAL CONDITIONS OF PURCHASE IS REQUIRED FOR COMPREHENSIVE GENERAL LIABILITY, AUTOMOBILE LIABILITY, AND WORKERS' COMPENSATION AND MUST BE SUBMITTED BY THE SUCCESSFUL BIDDER(S) TO THE DIVISION OF PURCHASES PRIOR TO AWARD. THE INSURANCE CERTIFICATE MUST NAME THE STATE OF RHODE ISLAND AS CERTIFICATE HOLDER AND AS AN ADDITIONAL INSURED. FAILURE TO COMPLY WITH THESE PROVISIONS MAY RESULT IN REJECTION OF THE OFFEROR'S BID. ANNUAL RENEWAL CERTIFICATES MUST BE SUBMITTED TO THE AGENCY IDENTIFIED ON THE PURCHASE ORDER. FAILURE TO DO SO MAY BE GROUNDS FOR CANCELLATION OF CONTRACT.

NOTE: IF THIS BID COVERS CONSTRUCTION, SCHOOL BUSING, HAZARDOUS WASTE, OR VESSEL OPERATION, APPLICABLE COVERAGES FROM THE FOLLOWING LIST MUST ALSO BE SUBMITTED TO THE DIVISION OF PURCHASES PRIOR TO AWARD: * PROFESSIONAL LIABILITY INSURANCE (AKA ERRORS & OMISSIONS) - \$1 MILLION OR 5% OF ESTIMATED PROJECT COST, WHICHEVER IS GREATER. * BUILDER'S RISK INSURANCE - COVERAGE EQUAL TO FACE AMOUNT OF CONTRACT FOR CONSTRUCTION. * SCHOOL BUSING - AUTO LIABILITY COVERAGE IN THE AMOUNT OF \$5 MILLION. * ENVIRONMENTAL IMPAIRMENT (AKA POLLUTION CONTROL) - \$1 MILLION OR 5% OF FACE AMOUNT OF CONTRACT, WHICHEVER IS GREATER. * VESSEL OPERATION - (MARINE OR AIRCRAFT) - PROTECTION & INDEMNITY COVERAGE REQUIRED IN THE AMOUNT OF \$1 MILLION.

AWARD

THE STATE, AT ITS SOLE DISCRETION, SHALL RESERVE THE RIGHT TO MAKE ONE OR MULTIPLE AWARDS FOR THIS REQUIREMENT AND/OR TO REJECT ANY OR ALL BIDS.

DELIVERY PER AGENCY

DELIVERY OF GOODS OR SERVICES AS REQUESTED BY AGENCY.

RIVIP INFO – BID SUBMISSION REQUIREMENTS

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer. When delivering offers in person to One Capitol Hill, vendors are advised to allow at least one hour additional time for clearance through security checkpoints.

SPECIFICATIONS: BLADDER SCANNER

Measurements: Type:	Bladder Volume
Patient Types: For Adults:	Yes
For Children:	No
Key Features: Screen Display:	Grayscale
Onboard Video Tutorial:	No
Onboard Printer:	Yes
Battery Charger/Wireless Hub:	Yes
Voice Annotation:	No
Cart-based* vs. Handheld (* optional):	Handheld or use with Cart
Technology for exam transfer to EHRs:	No
Range: Bladder Measurement:	0-999 ml
Bladder Volume: 700 to 999 ml +/-25%, 25 ml	0 to 699 ml +/-20%, +/-20 ml
Scan Time - Expressed In Seconds:	< 5 seconds
Dimensions: Console: Height: 7 cm (2.75 in) Length: 32 mm (11.25 in)	Width: 22.9 cm (9 in)
Battery Charger/Wireless Hub: Length: 11.8 cm (4.65 in)	Width: 79 mm (3.1 in)
Optional Cart: Overall height: ~76 cm (~30 in) Weight: 6.8 kg (15 lb)	Tray: 48 x 48 cm (19 x 19 in)
Weight: Unit w/batteries:	340 g (12 oz)
Power/Battery: Batteries Supplied:	2
Battery Type:	7.2v NIMH
Average Battery Life:	>6 hrs continuous
Display: Type:	Grayscale LCD

SPECIFICATIONS: BLADDER SCANNER

Tilting Screen for Positioning:	No
Ultrasound Output Parameters:	
Temporal Average Power:	1 m W maximum
Focal 20 dB Beam-Area:	1.4 cm ²
Transducer Dimension:	13 mm diameter
Working Frequency:	2 MHz
Peak Instantaneous Intensity:	14 W/ cm ² maximum
Pulse Repetition Frequency (Dome Probe):	180 pulses / second
Pulse Repetition Frequency (Flat Probe):	120 pulses / second
Operating Conditions:	
Temperature:	10° to 40° C (50° to 104° F)
Humidity:	30% to 75% non condensing
Atmospheric Pressure:	700-1060 hPa
Storage Conditions:	
Temperature:	-20° to 60° C (-4° to 140° F)
Humidity:	20% - 95% non-condensing
Atmospheric Pressure:	500-1060 hPa
Calibration:	
Instructions:	Requires annual calibration.
Warranty:	
With Purchase:	1 Year
Standards & Approvals:	
Patent Numbers:	4926871, 5235985
Certified ISO:	13485:2003

MUST INCLUDE:

1. Portable Rolling Cart
2. On site In-service Training for Bladder Scanner from Vendor
3. Bladder Scanner Training Pack/Video (Acute) (English)

SPECIFICATIONS: BLADDER SCANNER

