



STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS

Department of Administration
DIVISION OF PURCHASES
One Capitol Hill
Providence, RI 02908-5855

Tel: (401) 574-8100
Fax: (401) 574-8387
Website: www.purchasing.ri.gov

June 16, 2016

ADDENDUM NUMBER ONE

RFQ # 7550710

TITLE: Hydraulic Autopsy Tables, DOH

Closing Date and Time: 7/7/16 at 11:30 AM

Per the issuance of this ADDENDUM # (1), 10 pages, including this cover sheet.

Specification Change /Addition / Clarifications

Attached please find a copy of the Bid Specifications and a better copy of the photos.
NOTE: These are the exact same specifications and photos as those that were posted earlier.



Rhode Island Department of Health
Office of State Medical Examiner's Office
48 Orms Street
Providence RI 02904

Specification for two (2) Hydraulic Autopsy Table/Carts

- The state of Rhode Island is inquiring to purchase two (2) Hydraulic Autopsy tables/carts, with the following for each table:
- Carrier Chassis Top Plastic, with X-Ray side rails color "Cream".
- A hydraulic carrier base with frame for above table top.
- A cover frame to protect the hydraulic carrier.
- The frame on the carts must be manufactured to fit into the current locking system on the autopsy sinks/stations, per the attached photos for clarifications of dimensions and overall photos of the autopsy sink.

The top of the autopsy sink counter is 36" high from the floor, (Refer to image 8).

The cot frame needs to be less than 34" wide, with bars to lock into the table's cutout so the table can drain into the sink, (Refer to image 2).

The horizontal bar for the cot frame is to lock into has two cutouts for the frame, there are 3" and 31" from the right, (Refer to image 2).

The height of the locking bar from the floor is 23", (Refer to Image 5).

The front uprights of the cot frame needs to be 2" wide and 2" deep to lock into the sink, (Refer to images 3&4).

- Freight charges must be included in your unit pricing