

BID SOLICITATION



STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS
 CAPITOL HILL
 PROVIDENCE RI 02908

BID NUMBER: B05662
TITLE: BLOOD GAS ANALYZER
BID OPENING DATE AND TIME:
10/21/2005 1:40 PM

BUYER: LINDA ROCHE
 PHONE #: (401) 222 - 2142 ext. 118
 BLANKET PERIOD: 11/1/2005 - 10/31/2010

B MENTAL HEALTH, RETARDATION & HOSPITALS
I MHRH ELEANOR SLATER HOSPITAL
L MANUEL MATHIAS BLDG, LABORATORY
L PO BOX 8269
T CRANSTON RI 02920
O

S MENTAL HEALTH, RETARDATION & HOSPITALS
H MHRH ELEANOR SLATER HOSPITAL
I MANUEL MATHIAS BLDG, LABORATORY
P PO BOX 8269
T CRANSTON RI 02920
O

Requisition Number(s): R76H064696

Item	Class-Item	Quantity	Unit	Unit Price	Total
	<p>BLANKET REQUIREMENTS: 11/1/05 - 10/31/10</p> <p>BIDDING</p> <p>(a) A single price shall be quoted for each item against which a proposal is submitted. This price will be the maximum in effect during the agreement period. Any price decline at the manufacturer's level shall be reflected in a reduction of the agreement price to the State.</p> <p>(b) Quantities, if any, are estimated only. The agreement shall cover the actual quantities ordering during the period. Deliveries will be billed at the single, firm, awarded unit price quoted regardless of the quantities ordered.</p> <p>(c) Bid price is net F.O.B. destination and shall include inside delivery at no extra cost.</p> <p>(d) Bids for single items and/or a small percentage of total items listed, may, at the State's sole option, be rejected as being non-responsive to the intent of this request.</p> <p>ORDERING</p> <p>(a) The User Agency(s) will submit individual orders for the various items and various quantities as may be required during the agreement period.</p> <p>(b) Exception - Regardless of any agreement resulting from this bid, the State reserves the right to solicit prices separately for any extra large requirements for delivery to specific destinations.</p>				

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer. When delivering offers in person to One Capitol Hill, vendors are advised to allow at least one hour additional time for clearance through security checkpoints.

DELIVERY: _____

RIVIP VENDOR ID#: _____

TERMS OF PAYMENT: _____

**DO NOT SIGN BID ON THIS PAGE!
 USE CERTIFICATION COVER FORM.**

BID SOLICITATION



STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS
 CAPITOL HILL
 PROVIDENCE RI 02908

BID NUMBER: B05662
TITLE: BLOOD GAS ANALYZER
BID OPENING DATE AND TIME:
10/21/2005 1:40 PM

BUYER: LINDA ROCHE
 PHONE #: (401) 222 - 2142 ext. 118
 BLANKET PERIOD: 11/1/2005 - 10/31/2010

B MENTAL HEALTH, RETARDATION & HOSPITALS
I MHRH ELEANOR SLATER HOSPITAL
L MANUEL MATHIAS BLDG, LABORATORY
L PO BOX 8269
T CRANSTON RI 02920
O

S MENTAL HEALTH, RETARDATION & HOSPITALS
H MHRH ELEANOR SLATER HOSPITAL
I MANUEL MATHIAS BLDG, LABORATORY
P PO BOX 8269
T CRANSTON RI 02920
O

Requisition Number(s): R76H064696

Item	Class-Item	Quantity	Unit	Unit Price	Total
1.0	<p>THIS IS A MULTI-YEAR BID/CONTRACT. PER RHODE ISLAND STATE LAW 37-2-33, CONTRACT OBLIGATIONS BEYOND THE CURRENT FISCAL YEAR ARE SUBJECT TO AVAILABILITY OF FUNDS. CONTINUATION OF THE CONTRACT BEYOND THE INITIAL FISCAL YEAR WILL BE AT THE DISCRETION OF THE STATE. TERMINATION MAY BE EFFECTED BY THE STATE BASED UPON DETERMINING FACTORS SUCH AS UNSATISFACTORY PERFORMANCE OR THE DETERMINATION BY THE STATE TO DISCONTINUE THE GOODS/SERVICES, OR TO REVISE THE SCOPE AND NEED FOR THE TYPE OF GOODS/SERVICES; ALSO MANAGEMENT OWNER DETERMINATIONS THAT MAY PRECLUDE THE NEED FOR GOODS/SERVICES.</p> <p>979-53 POINT OF CARE BLOOD GAS ANALYZER REAGENT RENTAL COST PER TEST</p> <p>MEASURED PARAMETERS INCLUDE: pH, pCO2, pO2, tHb, sO2, FO2hb, FCOhb, FHHb, and FMetHb</p>	60.00	MO		

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer. When delivering offers in person to One Capitol Hill, vendors are advised to allow at least one hour additional time for clearance through security checkpoints.

DELIVERY: _____

RIVIP VENDOR ID#: _____

TERMS OF PAYMENT: _____

**DO NOT SIGN BID ON THIS PAGE!
 USE CERTIFICATION COVER FORM.**

BID SOLICITATION



STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS
 CAPITOL HILL
 PROVIDENCE RI 02908

BID NUMBER: B05662
TITLE: BLOOD GAS ANALYZER
BID OPENING DATE AND TIME:
10/21/2005 1:40 PM

BUYER: LINDA ROCHE
 PHONE #: (401) 222 - 2142 ext. 118
 BLANKET PERIOD: 11/1/2005 - 10/31/2010

**B
I
L
L
T
O** MENTAL HEALTH, RETARDATION & HOSPITALS
 MHRH ELEANOR SLATER HOSPITAL
 MANUEL MATHIAS BLDG, LABORATORY
 PO BOX 8269
 CRANSTON RI 02920

**S
H
I
P
T
O** MENTAL HEALTH, RETARDATION & HOSPITALS
 MHRH ELEANOR SLATER HOSPITAL
 MANUEL MATHIAS BLDG, LABORATORY
 PO BOX 8269
 CRANSTON RI 02920

Requisition Number(s): R76H064696

Item	Class-Item	Quantity	Unit	Unit Price	Total
	1) FIVE YEAR CONTRACT, REAGENT BASED ON COST PER TEST. 2) BASED ON AN AVERAGE OF FIVE TESTS PER DAY, AND TWENTY-FOUR CONTROLS PER YEAR AND TWO LINEARITY KITS ANNUALLY. 3) TRADE IN EXISTING INSTRUMENT, IRMA SYSTEM. BUY OUT AT THE END OF THE AGREEMENT PERIOD. THE SYSTEM MAY BE PURCHASED FOR ONE DOLLAR. 4) FIVE YEAR WARRANTY SERVICE AGREEMENT (MONDAY-FRIDAY 8:00 A.M. TO 5:00 P.M.) TO INCLUDE UNLIMITED ON DEMAND ON SITE SERVICE VISITS COVERING SERVICE PARTS, LABOR, TRAVEL AND ON PREVENTATIVE MAINTENANCE VISIT ANNUALLY TO BEGIN IN YEAR TWO. 5) THOROUGH INSTALLATION ON PREMISES, IN SERVICE AND USER TRAINING. 6) INSTRUMENT MUST BE NEW, NOT USED OR REFURBISHED.				

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer. When delivering offers in person to One Capitol Hill, vendors are advised to allow at least one hour additional time for clearance through security checkpoints.

DELIVERY: _____

RIVIP VENDOR ID#: _____

TERMS OF PAYMENT: _____

**DO NOT SIGN BID ON THIS PAGE!
 USE CERTIFICATION COVER FORM.**

BID SOLICITATION



STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS
 CAPITOL HILL
 PROVIDENCE RI 02908

BID NUMBER: B05662
TITLE: BLOOD GAS ANALYZER
BID OPENING DATE AND TIME:
10/21/2005 1:40 PM

BUYER: LINDA ROCHE
 PHONE #: (401) 222 - 2142 ext. 118
 BLANKET PERIOD: 11/1/2005 - 10/31/2010

**B
I
L
L
T
O** MENTAL HEALTH, RETARDATION & HOSPITALS
 MHRH ELEANOR SLATER HOSPITAL
 MANUEL MATHIAS BLDG, LABORATORY
 PO BOX 8269
 CRANSTON RI 02920

**S
H
I
P
T
O** MENTAL HEALTH, RETARDATION & HOSPITALS
 MHRH ELEANOR SLATER HOSPITAL
 MANUEL MATHIAS BLDG, LABORATORY
 PO BOX 8269
 CRANSTON RI 02920

Requisition Number(s): R76H064696

Item	Class-Item	Quantity	Unit	Unit Price	Total
	BLOOD GAS ANALYZER SPECIFICATIONS 1) MEASURE pH, pCO2, AND pO2 WITH A SINGLE USE CUVETTE. 2) USE ONLY 90 uL OF WHOLE BLOOD IN APPROXIMATELY ONE MINUTE. 3) DRI-TEK MEASURING TECHNOLOGY THAT DOES NOT REQUIRE CARTRIDGE PREPARATION OR MAINTENANCE. 4) SAMPLE IS ASPIRATED AUTOMATICALLY WITHOUT OPERATOR INTERVENTION. 5) DRI-TEK CARTRIDGES WITH A TWO YEAR SHELF LIFE AT ROOM TEMPERATURE AND THREE MONTH INSTALLED IN THE ANALYZER. 6) NO CONTACT WITH BIOHAZARDS, CARTRIDGES ARE SEALED FOR SAFE DISPOSAL. 7) NEW CUVETTE WITH INLET FOR EVERY SAMPLE. 8) BUILT IN ELECTRONIC QC SYSTEM, CHECKS THE EVERY CUVETTE BEFORE MEASUREMENT, RUN A TWO LEVEL CHECK FOR EACH MEASUREMENT.				

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer. When delivering offers in person to One Capitol Hill, vendors are advised to allow at least one hour additional time for clearance through security checkpoints.

DELIVERY: _____

RIVIP VENDOR ID#: _____

TERMS OF PAYMENT: _____

**DO NOT SIGN BID ON THIS PAGE!
 USE CERTIFICATION COVER FORM.**

BID SOLICITATION



STATE OF RHODE ISLAND AND PROVIDENCE PLANTATIONS
 CAPITOL HILL
 PROVIDENCE RI 02908

BID NUMBER: B05662
TITLE: BLOOD GAS ANALYZER
BID OPENING DATE AND TIME:
10/21/2005 1:40 PM

BUYER: LINDA ROCHE
 PHONE #: (401) 222 - 2142 ext. 118
 BLANKET PERIOD: 11/1/2005 - 10/31/2010

B MENTAL HEALTH, RETARDATION & HOSPITALS
I MHRH ELEANOR SLATER HOSPITAL
L MANUEL MATHIAS BLDG, LABORATORY
L PO BOX 8269
T CRANSTON RI 02920
O

S MENTAL HEALTH, RETARDATION & HOSPITALS
H MHRH ELEANOR SLATER HOSPITAL
I MANUEL MATHIAS BLDG, LABORATORY
P PO BOX 8269
T CRANSTON RI 02920
O

Requisition Number(s): R76H064696

Item	Class-Item	Quantity	Unit	Unit Price	Total
	9) CHECKS EVERY SAMPLE FOR CLOTS AND BUBBLES. 10) EFFICIENT DATA MANAGEMENT SYSTEM, WITH CAPACITY FOR 1000 PATIENT AND QC RESULTS. 11) FLEXIBLE SOFTWARE SYSTEM WITH PASSWORD PROTECTION AND COMPREHENSIVE SYSTEM OF REMOTE MONITORING. 12) POWER SOURCE AC AND RECHARGEABLE BATTERY. BRAND _____ MODEL # _____ DELIVERY OF GOODS OR SERVICES AS REQUESTED BY AGENCY.				
				TOTAL:	_____

It is the Vendor's responsibility to check and download any and all addenda from the RIVIP. This offer may not be considered unless a signed RIVIP generated Bidder Certification Cover Form is attached and the Unit Price column is completed. The signed Certification Cover Form must be attached to the front of the offer. When delivering offers in person to One Capitol Hill, vendors are advised to allow at least one hour additional time for clearance through security checkpoints.

DELIVERY: _____

RIVIP VENDOR ID#: _____

TERMS OF PAYMENT: _____

**DO NOT SIGN BID ON THIS PAGE!
 USE CERTIFICATION COVER FORM.**

BID FOR THE BLOOD GAS ANALYZER

NPT7 (Radiometer America Inc. 810 Sharon Drive, Westlake Ohio 44145)

- 1) Five year contract, reagent rental based on cost per tests.
- 2) Based on an average of five tests per day, and twenty- four controls per year, and two linearity kits annually.
- 3) Trade in of existing instrument the IRMA system. Buy out at the end of the agreement period, the system may be purchased for one dollar.
- 4) Five year warranty service agreement (Monday-Friday 8:00 to 5:00 pm), to include unlimited on demand on site service visits covering service parts, labor, travel and one preventative maintenance visit annually to begin in year two.
- 5) Thorough installation on premises, in service and user training.
- 6) Instrument must be new not used or refurbished.

BLOOD GAS ANALYZER SPECIFICATIONS

The system must meet or exceed the following requirements:

- 1) Measure pH, pCO₂, and pO₂ with a single use cuvette.
- 2) Use only 90 ul of whole blood, in approximately 1 minute.
- 3) Dri-tek measuring technology, that does not require cartridge preparation or maintenance.
- 4) Sample is aspirated automatically without operator intervention.
- 5) Dri- tek cartridges with a two year shelf life at room temperature and three month when installed in the analyzer
- 6) No contact with biohazards, cartridges are sealed for safe disposal.
- 7) New cuvette with inlet for every sample.

Eleanor Slater Hospital Laboratory
P.O. Box 8269
Cranston, Rhode Island

- 8) Built in electronic QC system, checks the every cuvette before measurement, run a two level check for each measurement.
- 9) Checks every sample for clots and bubbles.
- 10) Efficient data management system, with capacity for 1000 patient and QC results.
- 11) Flexible software system with password protection, and comprehensive system of remote monitoring.
- 12) Power source AC and rechargeable battery.