



**State of Rhode Island  
Department of Administration / Division of Purchases  
One Capitol Hill, Providence, Rhode Island 02908-5855  
Tel: (401) 574-8100 Fax: (401) 574-8387**

**Solicitation Information  
11/29/2019**

**ADDENDUM# 1**

**RFQ #7599857**

**TITLE: X-RAY TABLE – MEDICAL EXAMINER - RIDOH**

**Bid Closing Date & Time:** Wednesday, December 18, 2019 @ 10:00 AM  
Eastern Time (ET)

---

**Notice to Vendors**

**Attached are vendor questions with State responses. No further questions will be answered.**

---

**Anthony Venditelli  
Buyer I**

*Interested parties should monitor this website, on a regular basis, for any additional information that may be posted.*

**Vendor Questions with State Responses for RFQ #7599857 – X-RAY  
TABLE – MEDICAL EXAMINER - RIDOH**

Question 1: Would you be willing to consider a bid for a comparable/equivalent model radiographic table? The model we would be proposing has a weight capacity of 660lbs rather than 700lbs which has proven to be sufficient for all customers we have served so far. The potential advantage would be cost savings while still providing the same functionality.

Answer to question 1: No we have had combined weights (decedents, medical equipment or implements in or on decedents) over 660lb. We found the Del table tough enough to handle the abuse.

Question 2: In regard to RFQ 7599857, you have stated the x-ray table must be a new Del RT-100 4-Way Float Top Table. If [Insert Vendor Name] would be able to supply a table that meets/exceeds the required qualifications listed in the RFQ would we be included in consideration? Or is the request not open to any other manufacturers?

Answer to question 2: No substitutions, the table we now have is Del RT-100 4-way float top it is a well built, known product. We never had any problems with this make or model.